

Direction for use for pH/ Redox Combination Electrodes (KCl / Gel filled)

1. Remove the bottle/ protective cap at the bottom of the electrode.
2. Shake the electrode gently to ensure that the internal buffer solution covers the whole pH Bulb and no air bubbles are entrapped.
3. pH/ Redox Combination Electrodes should be filled with the appropriate electrolyte to a height of about 1cm below the filling port.
4. Wash off any salt film present on the exterior of pH/ Redox Combination Electrode, using distilled water.
5. To ensure pressure equalisation, the stopper of the filling port should be removed or perforated with a pin.
6. Soak the electrode in 7pH Buffer or saturated KCl for some hours (preferably overnight) before use.
7. Electrode which have developed problem of slow response due to drying out of the membrane or use under extreme conditions, may be re-activated by soaking in 0.1 N Hydrochloric Acid for 8 to 10 hours.

Electrode, which fails to respond to the above treatment can be further activated by dipping them to 2% Hydrofluoric Acid for 5-10 seconds and immediately washing with distilled water. This drastic treatment should however be applied only as last resort as it reduces the life of the electrode.

8. After conditioning the electrode as in (7) above, if it is not to be used it should be kept in 7pH Buffer or saturated KCl. To prevent the entry of measuring media through the diaphragm of the reference junction of pH Combination Electrode, the stopper of the filling opening/ port should be removed or perforated with a pin.

9(i) Calibration of pH Combination Electrode (KCl/ Gel filled) :

- a.) The pH Meter should have a capability of a minimum of Single point calibration for correcting asymmetry potential at 7pH.
- b.) The calibration of a pH measuring system is described in the Instruction Manual of the meter. For convenience it is repeated below:

The Electrode should be immersed in a buffer solution of known pH say 7pH. The display of the pH meter is brought to the value 7 by means of the asymmetry potential adjustment pot. The electrode is then removed from the buffer solution and washed with distilled water and immersed in another known buffer say 4pH. The display of the pH meter is brought to the value 4 by means of slope adjustment pot. The display of the measuring instruments should now indicate this value and if it does so, the electrode is functioning correctly.

The effect of temperature on the pH value of the buffer solution must be considered. The preceding

calibration procedure should be done initially and repeated at intervals determined by the desired measuring accuracy and the conditions of use of the electrode system.

9(ii) Calibration of Redox Combination Electrode (KCl/ Gel filled) :

- a.) The Redox Meter should have a capability of a minimum of Single point calibration for correcting offset.
- b.) Please follow instructions at Sr. No. 1, 3, 4 & 5 only and not others from "Direction for use".
- c.) Prepare a Redox Buffer solution for 264 mV & immerse the Redox Electrode in the same. After putting the meter in mV mode the display of the meter should be brought to 264 mV by means of the asymmetry potential adjustment pot. The electrode is then removed from the buffer solution and washed with distilled water and then immersed in another known buffer of 400 mV. If the reading now displayed is 400 ±25 mV then electrode is functioning correctly & ready for use.

Important for pH/ Redox Combination Electrodes (KCl / Gel filled):

- 10.a) Always keep the electrodes dipped in saturated KCl solution, when not in use.
- b.) Since the shelf life of the Redox buffer is very short, for correct results it is strongly recommended to use freshly prepared buffers for calibration.
- c.) Electrode should be kept in its box as per "arrow mark" when not in use, which means the bottle/ protective cap side should be facing the ground in its packed condition.

Guide Lines for Replacement

1. pH / Redox Electrodes are very fragile in nature and are tested before dispatch from the factory to ensure that they function correctly. Any electrode that fails to work for the first time when supplied on FORD prices and when used in accordance with the instructions provided in the "Direction for use", would be repaired / replaced at company's sole discretion. Failure of electrodes if any, should be informed within 30 days from the date of invoice, else no claim whatsoever will be entertained.
2. In any case, kindly do not take any step not mentioned in our "Direction for use" (Moreover, do not remove the cap on top of the electrode which will make the electrode ineffective).
3. It may kindly be noted that in case the above Guide Lines for Replacement is not followed, FREE REPLACEMENT will not be considered even if complaints are received within the period specified in (1) above.

TOSHNIWAL INSTRUMENTS MFG. PVT. LTD.,

P.O. Gagwana 305023, Dist. Ajmer (Raj.) INDIA

Phone: +91-145-2971131/ 2/ 3

Customer Care: +91-9829112345

Email: info@toshcon.com, Website: www.toshcon.com

REGD. OFFICE: 401, Manish Chamber, Sonawala Cross Lane, Goregaon (E), Mumbai-68